

Technical Manual
**APPLICATION DETAILS FOR THE RL
RANGE OF LOAD CELLS**

Doc. Ref CD1060H



Affirmed by Declaration
of Conformity

USA & Canada
RDP Electrosense Inc.

2216 Pottstown Pike
Pottstown, PA 19465
U.S.A.

Tel (610) 469-0850
Fax (610) 469-0852
E-mail info@rdpe.com
www.rdpe.com

All other countries
RDP Electronics Ltd

Grove Street, Heath Town,
Wolverhampton, WV10 0PY
United Kingdom

Tel: +44 (0) 1902 457512
Fax: +44 (0) 1902 452000
E-mail: sales@rdpe.com
www.rdpe.com

APPLICATION DETAILS FOR THE RL RANGE OF LOAD CELLS

For general handling and installation application notes refer to customer document CD1039.

The RL range comprises flat ‘pancake’ (compression and universal) load cells, and beam load cells for weighing and force measurement applications.

1. Summary of models in the RL range

Model	Type	Ranges (kg)	Material	Sealing	Nominal Bridge Resistance (ohms)
RLC	Compression	100 to 50,000	Tool steel	IP65	350/700
RLT	Tension	5 to 2,500	Tool steel	IP65	350
RLW	Weighing	10 to 1,500	Tool steel	IP67	350
RLS	Single point weighing	10 to 100	Aluminium Alloy	IP65	350
RLU	Tension/compression	100 to 10,000	Tool steel	IP65	350/700

- The materials used in the construction of the load cells are shown in the table above. As an option it is possible to replace the tool steel with stainless steel.
- The nominal bridge resistances given (see table above) are for the output (green-white cable cores). The tolerance is less than ± 10 ohm. The input resistance is slightly higher – typically 385 ± 30 ohms for nominal 350 ohm bridges and 750 ± 30 ohms for nominal 700 ohm bridge.
- The sensitivity (all models) is 2 mV/V FS $\pm 10\%$ eg at 10 volts excitation the full load output will be 20 mV $\pm 10\%$.
- The safe overload capacity is 150% FS (all models) and the ultimate overload capacity is 200% FS.
- Insulation resistance: $\geq 1000\text{M}\Omega$ at 50VDC (all models).
- Performance:

Parameter	Percentage of full range	
	Models U, C, T, W	Model S
Non-linearity (max)	$\pm 0.03\%$	$\pm 0.02\%$
Hysteresis (max)	$\pm 0.03\%$	$\pm 0.02\%$
Non-repeatability (max)	$\pm 0.03\%$	$\pm 0.02\%$

8. Environmental (all models):

Operating temperature range: -20 to +60°C
 Compensated temperature range: -10 to +40°C

Temperature Coefficients	Models		
	U,C	S	T, W
	% FS per °C	% FS per °C	% FS per °C
Zero (max)	±0.002	±0.0025	±0.003
Span (max)	±0.002	±0.0025	±0.003

9. Electrical connection:

Each load cell is fitted with an integral 3 metre long four core screened cable. The screen is not connected to the load cell body. The load cell body will normally be grounded via the mounting metalwork. The cable screen should be connected to the instrument.

The cable cores are:

Core Colour

Red excitation positive
 Black excitation negative
 Green output positive
 White output negative

Connection diagrams showing RL connections to various RDP instruments are:

<u>Diagram</u>	<u>Instrument</u>	<u>Diagram</u>	<u>Instrument</u>
D17031	E308	D17036	TR150
D17033	S7DC	D17035	611
D17034	S7MZ	D17037	E725-DC1
D17889	DR7DC		

10. Outline Drawings:

<u>Drawing</u>	<u>Load cell model</u>	<u>Drawing</u>	<u>Load cell model</u>
D16999	RLC	D17003	RLS
D17000	RLT	D16846	RLU
D17002	RLW		

11. Calibration

Each load cell is individually calibrated with equipment traceable to national standards, and an individual sensitivity figure is given.

RDP offers a calibration service when a load cell and instrument are purchased together. The instrument set-up to read the load/force in units specified by the customer.

WARRANTY AND SERVICE

WARRANTY.

R.D.P. Electronics products are warranted against defects in materials or workmanship. This warranty applies for one year from the date of delivery. We will repair or replace products that prove to be defective during the warranty period provided they are returned to R.D.P. Electronics.

This warranty is in lieu of all other warranties, expressed or implied, including the implied warranty of fitness for a particular purpose to the original purchaser or to any other person. R.D.P. Electronics shall not be liable for consequential damages of any kind.

If the instrument is to be returned to R.D.P. Electronics for repair under warranty, it is essential that the type and serial number be quoted, together with full details of any fault.

SERVICE.

We maintain comprehensive after-sales facilities and the instrument can, if necessary be returned to our factory for servicing.

Equipment returned to us for servicing, other than under warranty, must be accompanied by an official order as all repairs and investigations are subject to at least the minimum charge prevailing at the date of return.

The type and serial number of the instrument should always be quoted, together with full details of any fault and services required.

IMPORTANT NOTES.

1. No service work should be undertaken by the customer while the unit is under warranty except with the authorisation of RDP Electronics.
2. If the instrument is to be returned to R.D.P. Electronics for repair, (including repair under warranty) it is essential that it is suitably packed and that carriage is insured and prepaid. R.D.P. Electronics can accept no liability whatsoever for damage sustained during transit.
3. It is regretted that the above warranty only covers repairs carried out at our factory. Should the instrument have been incorporated into other equipment that requires our engineers to perform the repair on site, a charge will be made for the engineer's time to and from the site, plus any expenses incurred.

The aforementioned provisions do not extend the original warranty period of any product that has been either repaired or replaced by R.D.P. Electronics.

**THIS WARRANTY MAY BE NULL AND VOID SHOULD
THE CUSTOMER FAIL TO MEET OUR TERMS OF PAYMENT.**